



**Principal Investigator:**

Andrew Steele (R242, 202-478-8974)

**Secondary Contacts:**

Tim Strobel (R105, 202-478-8943)

George D. Cody (R218, 202-478-8980)

**Emergency Information:**

Staff Member: 301-204-3771 (cell)

BBR (Gary Bors): 202-510-8577

All other emergencies: 911

**Purpose:**

This laboratory is general laboratory use for chemical analysis, astro, micro, and molecular biology as well as sample preparation. It contains all the facilities for a Class 2 bio-safety rating but only Class 1 organisms should be used (see below).

*All laboratories and facilities on the Broad Branch Rd. campus are controlled areas. Specific training must be completed and documented before working in this laboratory / facility.*

**1. Chemicals Standard Operational Protocol**

*1.1 Chemical and Microbial Handling, during imaging and analysis, storage, and disposal:*

Study MSDA sheets on all chemicals that you work with (hard copy located in lab coat locker) or check on-line (<https://www.msdonline.com>) or with the product or kit manufacturer.

Microbes should be handled in biological safety cabinets. Only class I (class II on a specific case by case basis agreed with PI before ordering) biohazard is allowed in the laboratory. All class I and II organisms must be handled in biosafety hoods using aseptic technique. If you are not trained in aseptic technique please make that plain during safety walk through.

Appropriate protection must be worn when transferring chemicals, during sample preparation and when handling microbes and media, including lab coats, protective gloves and eye ware.

All media and culture slides, thin sections and plates etc must be clearly labeled with: media type, organism, date cultured and initials of investigator (label bottom of media plates not the lid).

All biologicals must be disposed of by autoclaving and bottles must be washed after autoclaving. Specialized media may need to be disposed of separately and must be agreed on with PI before use. Any used slides should be placed in a sharps box for disposal.

Any used media plates must be placed into autoclave bags for autoclaving. Autoclaved bags must be placed into Biohazard waste boxes for incineration. Autoclave (in R244) use for trained personnel only. Training will be given in a specialized course from the PI.

All cuts and abrasions must be treated using standard antiseptic and first aid measures and reported to the PI ASAP both in person and by email.

Date each chemical to record the day it was opened and by whom.

Only take what you need and do not return unused chemicals back to main reservoir. All chemicals must be stored in appropriate cabinets. Acids and Bases are never stored together.



Only dilute mineral acids may be discarded into marble chip baths located in the fume hoods. No chemicals may ever be poured into any sink or drain.

For biological spills use Bio-Spill kits, and all biological waste has to be autoclaved. Spill kits are located under the sink near the freezer.

Needles, pipette tips and sharp tools are disposed in suitable sharps protective boxes.

Chemicals that need to be disposed of shall be stored in appropriate cabinets until laboratory-wide storage removal is initiated (~ annually). When purchasing chemicals, try to buy as little as possible to minimize waste.

*1.2 Lab Chemicals: Where they are, how they are obtained for use:*

Familiarize yourself with where the various chemicals are stored based on type, reactivity, and flammability.

Familiarize yourself with the type of organisms being used in your experiments.

*1.3 Sample Labeling:*

All samples and media plates must be labeled in such a way as to be immediately identifiable (i.e. on the base not lid of plates). The use of notebook numbers or other schemes is not sufficient. Unlabeled vials constitute a serious offense and can lead to loss of laboratory privileges.

## **2. Instrument Standard Operational Protocol**

*Users must be familiar with the specific safety concerns of each instrument. Contact PI for details.*

*2.1 PCR: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab. **Warning:** Hot surfaces.*

*2.2 RT-PCR: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab.*

*2.3 Nanodrop: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab.*

*2.4 Array Printer: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab.*

*2.5 Centrifuge: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab. **Warning:** ensure all tubes are balanced before use.*

*2.6 BioAnalyzer: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab.*

*2.7 Biosafety Hoods: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab. **Warning:** danger of UV light exposure.*



2.7 *Cavity Ring Down Spectrometer*: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab. **Warning**: hidden lasers do not disassemble.

2.8 *Analytical Potentiostat*: Access to this instrument is subject to scheduling. All users must be trained or verified proficient by the PI first before using any instrument in the lab. **Warning**: electrocution danger.

### 3. Compressed Gasses, storage, and disposal

Appropriate regulators are required for compressed gasses.

3.1 *Storage of new Tanks*: All compressed tanks must be stored in the storage area. Unused tanks must have end-caps securely fastened. Do not store reducing and oxidizing gases in close proximity. Do not use carbon monoxide unless directly linked to fume hood.

3.2 *Securing in-use Tanks*: All in-use tanks must be securely attached to fixed bodies, e.g. bench top.

3.3 *Return of empty or long-term non-use tanks*: All gas and N<sub>2</sub> taps should be closed and checked before leaving the laboratory.

### 4. General Laboratory Safety Features: Lab Safety Rules:

4.1 *Eye-ware-protection*: Required for: All chemical reactions, microbial culturing, distillations, chemical transfer and sample preparation.

4.2 *Protective clothing*: Required for: All microbiology and chemical reactions, distillations, chemical transfer autoclaving Recommended for: sole-use operation of analytical instruments Cuts must be treated immediately by designated aiders.

### 5.0 Off-hours Operation:

5.1 *Any Chemistry*: Any chemistry performed in the hoods or on the bench that involve reactive compounds<sup>[SEP]</sup> (e.g. flammable liquids, metals, strong acids or bases, strongly exothermic reactions or the potential for explosion) should be performed during standard working hours to ensure that if an accident should occur proper response will be promptly initiated. In order to perform any chemical reactions that involve potentially reactive substances or products (see above) or have the potential for explosion (e.g. distillations or vacuum evaporation) after hours requires that a second person must be identified who is 1) in the building during such work 2) made clearly aware of the scope of the proposed work and the potential danger, and 3) is willing to check in at a frequency not less than once an hour to ensure safety. All bio spills should be cleaned immediately, locked and reported to lab supervisor ASAP.<sup>[SEP]</sup>

5.2 *Instruments*: The use of instruments after hours is not restricted but must be communicated with lab supervisor. The repair of instruments after hours is subject to the same restrictions as after-hours chemical processing.

6.0 **General Lab Safety Features**:<sup>[SEP]</sup> You will be made aware of the locations of; Fire extinguishers Emergency Shower, Eye wash station and Emergency Power Off Switch during laboratory safety training.



Keep Laboratory Doors Locked during off hours unless you are actively using the facility.

**Laboratory User**

*I agree that I have thoroughly read and understood this laboratory safety document. I have access to this safety information at all times when I am working. I have been trained to be able to identify the hazards to which I may be exposed and to follow the work practices and procedures discussed in this document. I certify that I will conduct my research work safely and that I will be responsible for following stated safety policies.*

\_\_\_\_\_  
User Name (Print)

\_\_\_\_\_  
User Signature

\_\_\_\_\_  
Date

**Principal Investigator**

*I certify that the information presented in this safety document is accurate and complete. I agree to comply with all safety procedures and to fully train and supervise all researchers under my direction.*

\_\_\_\_\_  
PI Signature

\_\_\_\_\_  
Date